



Telic, S.A.U.

Polígono Industrial Can Barri
C/ Moli d'en Barri, 7,
08415 Bigues i Riells, BARCELONA, Spain
Tel: +34 93 865 61 25
Fax: +34 93 865 62 46

EC DECLARATION OF CONFORMITY

TELIC, S.A.U. declares that the products listed in annexes of the present declaration have been manufactured according to requirements of the **Directive 93/42/EEC** and meet requirements set in the Essential Requirements of the Annex I of above mentioned Regulation.

Technical documentation, in accordance with the established in the corresponding annexes of the Directive 93/42/EEC, is updated and located in our facilities. We are in position to submit these documents in case of Notified Body or Competent Authority requirement.

This declaration applies to design, manufacturing and final control of medical devices. Validity of the present declaration is subject to the expiration of the corresponding EC certificates for different products.

Bigues i Riells, on March 25th 2021

Laura Delgado
Technical Manager

Oscar Lacruz
CEO



EC DECLARATION OF CONFORMITY – ANNEX 1
List of products with EC mark

Return electrodes for electrosurgery	
Description	Pre-gelled electrosurgical plates.
Commercial brand	BLAYCO
References	2125, 2125-5, 2125-C/00, 2125-C/00/5, 2125-C/10, 2125-C/10/5, 2125-C/21 2225, 2225-5, 2225-C/00, 2225-C/10 2425, 2425-C/00, 2425-C/10 2925, 2925-5, 2925-C/00, 2925-C/10 2500, 2500-5, 2500-C/00, 2500-C/12, 2500-C/12/5 2510, 2510-5, 2510-C/00, 2510-C/00/5, 2510-C/12 2600, 2600-C/00, 2600-C/12 2700, 2700-C/00, 2700-C/12 2900, 2900-5, 2900-C/00, 2900-C/12, 2900-C/12/5
Classification	
Product class IIb - Non-sterile. According to Rule 9 of Annex IX of the Directive 93/42/EEC	
GMDN	58494
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of the Directive 93/42/EEC	
Certificate number: 1434-MDD-250/2019	
Issued by: Polish Centre for Testing and Certification (PCBC)	
Notified Body number: 1434	
Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN 60601-2-2:2009//EN 60601-2-2:2009/A11:2011//EN ISO 15223-1:2012// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010//EN ISO 10993-5:2009//EN-ISO 10993- 10:2013// EN 60601-1:2006/EN 60601-1:2006/A1:2013 // EN 60601-2-2:2009// EN 60601-2- 2:2009/A11:2011. 2011/65/EU Directive (RoHS 2).	

Vein strippers	
Description	Single use vein strippers.
Commercial brand	DORMO-STRIP
References	VE-022, VE-025, VE-022 OP
Classification	
Product class IIa - Sterile. According to Rule 7 of Annex IX of the Directive 93/42/EEC	
GMDN	32321

**EC Full Quality Assurance System Certificate**

In accordance to Annex II (except Section 4) of the Directive 93/42/EEC

Certificate number: 1434-MDD-249/2019

Issued by: Polish Centre for Testing and Certification (PCBC)

Notified Body number: 1434

Valid until: 26/08/2022

Standards applied

EN ISO 14971:2012// EN 62366:2008/ EN 62366:2008/A1:2015// EN ISO 11607-1:2009/ EN ISO 11607-1:2009/A1:2014// EN ISO 11607-2:2006/ EN ISO 11607-2:2006/A1:2014// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC:2010// EN ISO 10993-4:2009// EN ISO 10993-5:2009// EN ISO 10993-7:2008/ EN ISO 10993-7:2008/AC:2009//EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN 556-1:2001/ EN 556-1:2001/AC:2006// EN ISO 11135-1:2007// CEN ISO/TS 11135-2:2008// EN ISO 14644-1:2015// EN ISO 14644-3:2005// EN ISO 14644-4:2001// EN ISO 14644-5:2004// EN ISO 14644-8:2013// EN ISO 15223-1:2012// EN 1041:2008+A1:2013.

Vascular loops

Description Two vascular loops for identification, occlusion, retraction.

Commercial brand DORMO-LOOP

References 405305, 402509, 402505, 402503, 402501, 401909, 401905, 401903, 401901, 400709

Classification

Product class IIa - Sterile. According to Rule 7 of Annex IX of the Directive 93/42/EEC

GMDN 61916

EC Full Quality Assurance System Certificate

In accordance to Annex V of the Directive 93/42/EEC

Certificate number: 1434-MDD-252/2019

Issued by: Polish Centre for Testing and Certification (PCBC)

Notified Body number 1434

Valid until: 26/08/2022

Standards applied

EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC:2009//EN ISO 10993-10:2013//EN ISO 10993-11:2009//// EN ISO 11135-1:2007// EN ISO 14644-1:2015//EN ISO 14644-2:2015//EN ISO 14644-4:2001//EN ISO 14644-5:2004// EN ISO 15223-1:2012.

Sterile ultrasound gel

Description Ultrasound transmission gel. Sterile.

Commercial brand TRANSONIC

References G-15E



Classification	
Product class I - Sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	58735
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of the Directive 93/42/EEC	
Certificate number: 1434-MDD-248/2019	
Issued by: Centre for Testing and Certification (PCBC)	
Notified Body number: 1434	
Valid until: 26/08/2022	

Standards applied
EN ISO 14971:2012// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010// EN ISO 10993-4:2009//EN ISO 10993-5:2009//EN ISO 10993-10:2013//EN ISO 10993-11:2009// EN ISO 15223-1:2012// EN ISO 11607-1:2009//EN ISO 11607-1:2009/A1:2014//EN ISO 11607-2:2006//EN ISO 11607-2:2006/A1:20//EN ISO 14644-1:2015//EN ISO 14644-3:2005//EN ISO 14644-4:2001//EN ISO 14644-5:2004//EN ISO 14644-8:2013 14// EN 556-1:2001//EN 556-1:2001/AC:2006//EN ISO 11137-1:2015//EN ISO 11137-2:2015// EN ISO 14644-1:2015//EN ISO 14644-3:2005//EN ISO 14644-4:2001//EN-ISO 14644-5:2004//EN ISO 14644-8:2013.

Cover for surgical light handle	
Description	Cover for surgical light handle.
Commercial brand	BLAYCO
References	LHC-01, LHC-03
Classification	
Product class I - Sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	44977
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of the Directive 93/42/EEC	
Certificate number: 1434-MDD-248/2019	
Issued by: Centre for Testing and Certification (PCBC)	
Notified Body number: 1434	
Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN 556-1:2001// EN 556-1:2001/AC: 2006// EN ISO 11135-1:2007// EN ISO 14644-1:2015// EN ISO 14644-3:2005// EN ISO 14644-4:2001// EN ISO 14644-5:2004// EN ISO 15223-1:2012.	



Tungsten electrosurgical accessories	
Description	Disposable sterile electrodes in tungsten
Commercial brand	BLAYCO
References	ABC-45, ABC-55, ABC-55/A30, ABC-55/A45, ABC-65
Classification	
Product class IIb - Sterile. According to Rule 9 of Annex IX of the Directive 93/42/EEC	
GMDN	61869
EC Full Quality Assurance System Certificate	
In accordance to Annex II, section 3 of the Directive 93/42/EEC	
Certificate number: 1434-MDD-251/2019	
Issued by: Centre for Testing and Certification (PCBC)	
Notified Body number: 1434	
Valid until: 10/05/2023	

Standards applied
ISO 13485:2003// EN 60601-1:2006// EN 60601-2-2:2009// EN ISO 14971:2012// EN ISO 10993-1:2009// EN 1041:2008// EN ISO 15223-1:2012// EN ISO 11135-1:2007// EN 556-1:2001/AC: 2006// EN ISO 11607-1:2009// EN ISO 11607-2:2006. 65/EU Directive (RoHS 2).

Electrode tip cleaner	
Description	Electrode tip cleaner.
Commercial brand	BLAYCO
References	AL-40
Classification	
Product class I - Sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	37483
EC Full Quality Assurance System Certificate	
In accordance to Annex V, section 3 of the Directive 93/42/EEC	
Certificate number: 1434-MDD-442/2019	
Issued by: Centre for Testing and Certification (PCBC)	
Notified Body number: 1434	
Valid until: 02/10/2021	
Standards applied	
EN ISO 14971:2012// EN 556-1:2001/ EN 556-1:2001/AC: 2006// EN ISO 11135-1:2007// EN 868-5:2009// ISO-2859-1:1999// EN ISO 15223-1:2012.	



EC DECLARATION OF CONFORMITY – ANNEX 2
List of self-certified products

Defibrillation electrodes without cable	
Description	Set of two adhesive pre-gelled pads with conductive hydrogel for defibrillation. To be used for adult patient use.
Commercial brand	DESFI-DORMO
References	ED-1010
Classification	
Product class I - Non-sterile. According to Rule 9 of Annex IX of the Directive 93/42/EEC	
GMDN	11130
Standards applied	
EN ISO 14971:2012// EN 60601-2-4:2011// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN-ISO 10993-5:2009// EN ISO 10993-7:2008// EN ISO 10993-7:2008/AC:2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012 Directiva 2011/65/EU (RoHS 2).	

Defibrillation electrodes with cable for adult patient	
Description	Set of two multi-function electrodes. Adult.
Commercial brand	DESFI-DORMO
References	EDC-1011, EDC-1015, EDC-1020, EDC-1025, EDC-1030, EDC-1035, EDC-1040, EDC-1045, EDC-1050, EDC-1055, EDC-1060, EDC-1065, EDC-1070
Description	Set of two multi-function electrodes. Adult. Pre-connected.
Commercial brand	DESFI-DORMO
References	EDC-2015, EDC-2020, EDC-2025, EDC-2030, EDC-2035, EDC-2035L, EDC-2040, EDC-2045, EDC-2050, EDC-2055, EDC-2060, EDC-2065, EDC-2070, EDC-2075, EDC-2080, EDC-2085, EDC-2090.
Classification	
Product class I – Non-sterile. According to Rule 9 of Annex IX of the Directive 93/42/EEC	
GMDN	45806
Standards applied	
EN ISO 14971:2012// EN 60601-2-4:2011// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN ISO 10993-5:2009// EN ISO 10993-7:2008// EN ISO 10993-7:2008/AC:2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012 Directiva 2011/65/EU (RoHS 2).	



Defibrillation electrodes with cable paediatrics	
Description	Set of two defibrillation electrodes. Paediatrics.
Commercial brand	DESFI-DORMO
References	EDC-P111, EDC-P115, EDC-P120, EDC-P125, EDC-P130, EDC-P135, EDC-P140, EDC-P145, EDC-P150, EDC-P155, EDC-P160, EDC-P165, EDC-P170, EDC-P175.
Description	Set of two defibrillation electrodes. Paediatrics. Pre-connected.
Commercial brand	DESFI-DORMO
References	EDC-P215, EDC-P220, EDC-P225, EDC-P230, EDC-P235, EDC-P240, EDC-P245, EDC-P250, EDC-P255, EDC-P260, EDC-P265, EDC-P270, EDC-P275, EDC-P280, EDC-P290.
Classification	
Product class I – Non-sterile. According to Rule 9 of Annex IX of the Directive 93/42/EEC	
GMDN	41587
Standards applied	
EN ISO 14971:2012// EN 60601-2-4:2011// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN ISO 10993-5:2009// EN ISO 10993-7:2008// EN ISO 10993-7:2008/AC:2009//EN ISO 10993-10:2013//EN ISO 10993-11:2009// EN ISO 15223-1:2012. ANSI/AAMI DF80:2003 2011/65/EU Directive (RoHS 2).	

ECG electrodes and accessories	
Description	ECG electrodes Ag/AgCl.
Commercial brand	DORMO
References	LEH-36, LF-50, LF-50T, LF-36, LP-50, LR-50, SX-50, SX-36, SF-36, SX-30, SP-50
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	35035
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC : 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC : 2009// EN ISO 10993-10:2013. ANSI/AAMI EC12 Disposable ECG electrodes. 2011/65/EU Directive (RoHS 2).	



Neonatal ECG electrodes	
Description	Neonatal ECG electrodes Ag/AgCl.
Commercial brand	DORMO
References	K-140, KS-140, K-150, KS-150, KF-140, KFS-140, KF-150, KFS-150, EKF-22KT, P-40/80 (non pre-gelled)
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	17460
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC : 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC : 2009// EN ISO 10993-10:2013. ANSI/AAMI EC12 Disposable ECG electrodes. 2011/65/EU Directive (RoHS 2)	

Resting electrodes and accessories	
Description	Resting electrodes.
Commercial brand	DORMO-TAB
References	T-2226
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	35035
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC: 2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012. ANSI/AAMI EC12 Disposable ECG electrodes. 2011/65/EU Directive (RoHS 2)	



TENS electrodes and recharges	
Description	Pre-gelled electrode for electrical stimulation.
Commercial brand	DORMO-TENS
References	DT-30R, DT-50R, DT-30, DT-50, DT-100 RT-30R, RT-50R, RT-30, RT-50, RT-100 T-1005, T-5055, ST-50, ST-100, ST-30R, ST-50R Conductive Silicone Strip CSC-1, CSC-25
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	35995

Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC: 2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012. ANSI/AAMI EC12:2000 2011/65/EU Directive (RoHS 2)	

Reusable cables for electrosurgery	
Description	Reusable clamp-cables for electrosurgical plates.
Commercial brand	BLAYCO
References	4200, 4200-5, 4210, 4210-5, 4212, 4212-5
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	47487
Standards applied	
EN ISO 14971:2012// EN 60601-2-2:2009/ EN 60601-2-2:2009/A11:2011// EN ISO 15223-1:2012. 2011/65/EU Directive (RoHS 2).	

Bite-blocks	
Description	Bite block for endotracheal tubes and laryngeal masks.
Commercial brand	MORDEDOR-MO
References	7600, 7650
Classification	
Product class I – Non-sterile. According to Rule 5 of Annex IX of the Directive 93/42/EEC	
GMDN	10405



Standards applied
EN ISO 9001:2015// EN 46002:1996.

Otoscope speculum	
Description	Disposable speculum for otoscope.
Commercial brand	DORMO-SPEC
References	4010, 4020, 4030, 4040, 4050, 4060, 4070, 4080, 4090, 4095
Classification	
Product class I – Non-sterile. According to Rule 5 of Annex IX of the Directive 93/42/EEC	
GMDN	35348
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012.	

Protective pad	
Description	Protective pad for surgical interventions.
Commercial brand	BLAYCO-PAD
References	AC-3020
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	62789
Standards applied	
EN ISO 14971:2012// EN 556-1:2001/ EN 556-1:2001/AC: 2006// EN ISO 11135-1:2007// EN ISO 14644-1:2015// EN ISO 14644-3:2005// EN ISO 14644-4:2001// EN ISO 14644-5:2004// EN ISO 15223-1:2012.	

Nasal holder for gastric catheters	
Description	Nasal holder for gastric catheters.
Commercial brand	DORMO-NAS
References	7500, 7525, 7550
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	62581
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012.	



Cold/hot packs	
Description	Reusable pack for Cold/Hot.
Commercial brand	DORMO, OXD
References	FC-01, FC-02, FC-03
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	37240
Standards applied	
EN-ISO 14971:2012// EN ISO 15223-1:2012.	

Ultrasound gels	
Description	Ultrasound transmission gel.
Commercial brand	TRANSONIC GEL, OXD BLUE
References	G-15, G-15/05, G-15/1, G-15/5, G-15/5RB, G-15A US-B250, US-B1, US-B5F, US-B5R
Description	Ultrasound transmission gel.
Commercial brand	TRANSONIC GEL CLEAR, OXD CLEAR
References	GC-15, GC-15/05, GC-15/1, GC-15/5, GC-15/5RB US-C250, US-C1, US-C5F, US-C5R
Classification	
Product class I – Non-sterile. According to Rule 5 of Annex IX of the Directive 93/42/EEC	
GMDN	15321
Standards applied	
EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-4:2009// EN ISO 10993-5:2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012.	

ECG Gel	
Description	Conductive gel for electrodes.
Commercial brand	ELECTRO-GEL
References	G-10, G-10A
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC	
GMDN	11425

**Standards applied**

EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-4:2009// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC: 2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012.
ANSI/AAMI EC12:2000.

Lubricating gel

Description	Lubricating water-soluble gel
Commercial brand	DORMO
References	G-20, G-20/5RB

Classification

Product class I – Non-sterile. According to Rule 5 of Annex IX of the Directive 93/42/EEC
GMDN 33587

Standards applied

EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012.

**EC DECLARATION OF CONFORMITY – ANNEX 3****Article 12 Directive 93/42/EEC Particular procedure for systems and procedure Packs**

In accordance with Article 12 Directive 93/42/EEC for Procedure Packs of articles with their own CE certificate.

Electrosurgical pencil with electrode tip cleaner	
Description	Electrosurgical pencil, hand control, with 70mm blade electrode and Electrode tip cleaner.
Commercial brand	BLAYCO
References	MB-200
Classification	
• <i>Electrosurgical pencil:</i> Product class IIb – Sterile. According to Rule 9 of Annex IX of the Directive 93/42/EEC • <i>Electrode tip cleaner:</i> Product class I – Sterile. According to Rule 1 of Annex IX of the Directive 93/42/EEC • Pack MB-200 Non-sterile.	
GMDN	Electrosurgical pencil: 61869 Electrode tip cleaner: 37483
Standards applied	
EN ISO 14971:2012// EN 60601-1:2006/EN 60601-1:2006/A1:2013// EN 60601-2-2:2009//EN 60601-2-2:2009/A11:2011//EN ISO 15223-1:2012. 2011/65/EU Directive (RoHS 2).	